

471 Simple Argon Triple Point Apparatus

PROVISIONAL DATA*

- Affordable
- Robust and simple to use
- Accurate to ±1mK 4 hour plateau typical

The Isotech Argon Triple Point Apparatus is a robust, simple to use and affordable solution for the realisation of the argon triple point.

Many laboratories use liquid nitrogen comparators which are convenient and can be low cost but the nitrogen boiling point is not on the ITS-90. More seriously the LN point is below that of Argon. Many standard platinum resistance thermometers (SPRTs) are filled with a mixture of argon and oxygen and at -195°C will be under a partial vacuum which affects the self-heating of the SPRT leading to a larger calibration uncertainty.

For many laboratories the high cost and complexity of previously available argon systems has been a barrier.

Now after years of research lsotech have introduced a new system that is more affordable, simple to use and will allow more laboratories the benefits of being able realise the argon triple point.

The Isotech system requires no electricity; the only consumable is liquid nitrogen - the 6N Pure argon is contained in a pressure vessel. A re-entrant tube allows liquid nitrogen to initially cool this volume to approximately -195°C. Weights are then added to a pressure release valve to increase the nitrogen's boiling temperature to just above the argon cells triple point.

The argon settles into its triple point for around four hours, allowing an SPRT inside the re-entrant tube to be calibrated.

*Provisional Data

The specification on this product is provisional, please visit the website for the latest information. http://www.isotech.co.uk/argon









Sectional view from front

Sectional view from side

ISOTHERMAL TECHNOLOGY LTD							Date of issue				
UKAS LABORATORY Nº 0175									1		
		Premium SPR1	Calibration	- Argor	TP			Author	sed by	0	
		Budget Nº.	1							1	
Note number (below)	Source	of uncertainty	Value ±	Unit	Probability distribution	Divisor	Sensitivity Ci	Standard Encertainty	Degrees of freedom y, or y,	u, ²	u ⁴ fv _i
1 2	Celibration of Sta Uncorrected Drift	ndard Rasistor since last calibration	0.000002	000	normal rectangular	1.73	1	0.000001	1	0.000000000	0
4	microK linearity	t	0.000007	c	normal	2.00	1	0.000004	1	0.000000000	0
<u>6</u> 7	Uncertainty of the Slope of cell melt	Fixed Point Cell during cal	0.001000	C C	normal rectangular	2.00	1	0.000500 0.000577	· · · · · · · · · · · · · · · · · · ·	D.000000250 D.000000333	0
8	Self Heating Effe	tainty cts	0.000500	C C	normal rectangular	1.00	1	0.000500	Ī	0.000000250	0 Q
10	Benestability of the	rostatic correction	0.000010	0	rectangular.	1.00	1	0.000010	16	0.000000000	0.35-22
13	Propagation of the	e water triple point u/c	0.000007	c	rectangular	1.73	1	0.000123	47353	0.000000000	0
ũ	Expanded uncert	ainty		_	normal	k =	2.00	0.001844	47353	0.0000000000	140-11
					×*************************************	ł	(=	2	2.00	0.00	1844
									-		
5.460 ► 5.455 ► 5.450 ► 5.445 ► 5.440 ► 5.435 ► 5.430 ► 0			A Time (hours)		10m 6	IK I	8				